

MADE IN FRANCE 

Regulatory information

CE Marking	Index of classification	PHTHALATES	LATEX
CE	I		

This device is intended to be used only once, but its design is very reliable but adapted for a single use.


1 – Device identification

Illustration of the assembly :



2 - References

Reference	Designation : Cuff with 1 bladder and 2 tubes	Color of tube	Dimensions in cm		Overall length
			Width	Length	
GBDU101Z	Cuff new born	White	3.5	29	38
GBDU102Z	Cuff child	Pink	4.5	35	48
GBDU103Z	Cuff small adult	Parma	6	46	61
GBDU104Z	Cuff adult	Blue sky	8	46	67
GBCU105Z	Conical cuff adult large	Turquoise	10	47	70
GCDU103Z	Lower limb cuff adult straight	Yellow	8	76	97
GCDU104Z	Lower limb cuff adult straight	Purple	10	62	76
GCDU105Z	Lower limb cuff adult straight L	Green	10	76	90
GCDU106Z	Lower limb cuff adult straight XL	Grey	10	82	100
GCDU107Z	Lower limb cuff adult straight XXL	Red	10	107	122
GCCU104Z	Conical lower limb cuff adult	Purple	10	62	76
GCCU105Z	Conical lower limb cuff adult L	Green	10	76	90
GCCU106Z	Conical lower limb cuff adult XL	Grey	10	82	100
GCCU107Z	Conical lower limb cuff adult XXL	Red	10	107	122



3 – Use of the device – Operating instructions

- Exsanguinate the limb using an Esmarch bandage or by elevating at 45° for 5 to 8 min.
- Before fitting the tourniquet:
 - check that the tourniquet to be fitted is of the correct size for the limb
 - check that the pocket of the tourniquet is sealed
 - check the quality of the Velcro and seams on the arm cuff / lower limb cuff
- Position the arm cuff or lower limb cuff on a protective bandage, previously applied on the skin of the limb on which the tourniquet is to be fitted.
- Insert the connecting tube fully (click) in the quick coupler.
- Before applying pressure, pull gently on the tube to check for perfect connection of the 2 parts.
- This device is intended to be used only once.

4 - Identification and characteristics of different subsets

Reference	Designation	Characteristics
	Arm cuff and lower limb cuff monobloc	Fabric coatings TPU - Latex free
A11007	Buckle	ACETAL
	Connecting tube	Medical PVC

5 - Cleaning and disinfection

Arm cuff and lower limb cuff are not intended to be reused :there is no protocol for cleaning and disinfection.

6 - Packaging

Reference	Packaging unit
GBDU101Z	10
GBDU102Z	10
GBDU103Z	10
GBDU104Z	10
GBCU105Z	5
GCDU103Z	5
GCDU104Z	5
GCDU105Z	5
GCDU106Z	5
GCDU107Z	5
GCCU104Z	5
GCCU105Z	5
GCCU106Z	5
GCCU107Z	5

7 - Storage

Type of packaging	Storage area	Temperature	Humidity	Atmospheric pressure
Original packaging	Medically clean	0° à 30 ° c	30 to 40 %	500 to 1060 hpA

8 - Warranty

This warranty provides assurance for the customer who purchases a D & D product that should the product fail to function to D & D published specifications during the term of this warranty, will either replace or repair.

The product must be used in accordance with its labeling and may not be altered or subjected to misuse, abuse, accident or improper handling.

Dessillons & Dutrillaux

Zone Industrielle La Tuque
Tél : +33 5 53 48 30 66
technique@ddmedical.fr

47240 CASTELCULIER
Fax : +33 5 53 47 24 64
www.ddmedical.fr

Date of first EC marking certificate :
2012

Page 4 / 4